



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Instrument Service & Controls

PO Box 55, Mendon, UT 84325

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Electrical, Thermodynamic, and Time and Frequency Calibration ***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

May 26, 2022

Issue Date:

July 16, 2024

Expiration Date:

September 30, 2026

Accreditation No.:

114450

Certificate No.:

L24-541

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pilabs.com



Certificate of Accreditation: Supplement

Instrument Service & Controls

PO Box 55, Mendon, UT 84325

Contact Name: Mr. Joseph Weeks Phone: 801 916-3320

Accreditation is granted to the facility to perform the following testing:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration Indication and Control Equipment used with Thermocouple - Type K ^o	-50 ^o C to 1 100 ^o C	0.5 ^o C	Temperature Simulator 33K5-4-586-1/	AMS2750
Equipment to Output DC mV ^o	-90 mV to 90 mV	0.015 mV	Altek 422 Temperature Simulator	AMS2750/OEM Manual

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Controllers with Thermocouple Type K ^o	30 ^o C to 1 149 ^o C	0.5 ^o C	Temperature Simulator Type K Thermocouple	AMS2750/OEM Manual
Temperature Controllers with Thermocouple Type T ^o	Up to 400 ^o C -80 ^o C to 0 ^o C	0.26 ^o C 0.6 ^o C	Temperature Simulator Type T Thermocouple	
Ovens, Chambers, Temperature Uniformity Survey ^o	-50 ^o C to 1 300 ^o C	0.9 ^o C	Temperature Indicator with Thermocouple Type K	AMS2750/SOP-OS-001

Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Timers/Recorders ^o	Up to 24 hrs	0.6 s/Day	Stopwatch	SP960-12/AMS2750/SOP-OS-001

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.



Certificate of Accreditation: Supplement

Instrument Service & Controls

PO Box 55, Mendon, UT 84325

Contact Name: Mr. Joseph Weeks Phone: 801 916-3320

Accreditation is granted to the facility to perform the following testing:

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
5. The term T represents temperature in °C or °F as appropriate to the uncertainty statement.

